



January 15, 2026

Curiteva, Inc.
Eric Linder
Chief Technology Officer
25127 Will Mccomb Dr.
Tanner, Alabama 35671

Re: K254061

Trade/Device Name: Curiteva Porous PEEK Cervical Interbody Fusion System
Regulation Number: 21 CFR 888.3080
Regulation Name: Intervertebral Body Fusion Device
Regulatory Class: Class II
Product Code: ODP
Dated: December 17, 2025
Received: December 17, 2025

Dear Eric Linder:

We have reviewed your section 510(k) premarket notification of intent to market the device referenced above and have determined the device is substantially equivalent (for the indications for use stated in the enclosure) to legally marketed predicate devices marketed in interstate commerce prior to May 28, 1976, the enactment date of the Medical Device Amendments, or to devices that have been reclassified in accordance with the provisions of the Federal Food, Drug, and Cosmetic Act (the Act) that do not require approval of a premarket approval application (PMA). You may, therefore, market the device, subject to the general controls provisions of the Act. Although this letter refers to your product as a device, please be aware that some cleared products may instead be combination products. The 510(k) Premarket Notification Database available at <https://www.accessdata.fda.gov/scripts/cdrh/cfdocs/cfpmn/pmn.cfm> identifies combination product submissions. The general controls provisions of the Act include requirements for annual registration, listing of devices, good manufacturing practice, labeling, and prohibitions against misbranding and adulteration. Please note: CDRH does not evaluate information related to contract liability warranties. We remind you, however, that device labeling must be truthful and not misleading.

If your device is classified (see above) into either class II (Special Controls) or class III (PMA), it may be subject to additional controls. Existing major regulations affecting your device can be found in the Code of Federal Regulations, Title 21, Parts 800 to 898. In addition, FDA may publish further announcements concerning your device in the Federal Register.

Additional information about changes that may require a new premarket notification are provided in the FDA guidance documents entitled "Deciding When to Submit a 510(k) for a Change to an Existing Device"

(<https://www.fda.gov/media/99812/download>) and "Deciding When to Submit a 510(k) for a Software Change to an Existing Device" (<https://www.fda.gov/media/99785/download>).

Your device is also subject to, among other requirements, the Quality System (QS) regulation (21 CFR Part 820), which includes, but is not limited to, 21 CFR 820.30, Design controls; 21 CFR 820.90, Nonconforming product; and 21 CFR 820.100, Corrective and preventive action. Please note that regardless of whether a change requires premarket review, the QS regulation requires device manufacturers to review and approve changes to device design and production (21 CFR 820.30 and 21 CFR 820.70) and document changes and approvals in the device master record (21 CFR 820.181).

Please be advised that FDA's issuance of a substantial equivalence determination does not mean that FDA has made a determination that your device complies with other requirements of the Act or any Federal statutes and regulations administered by other Federal agencies. You must comply with all the Act's requirements, including, but not limited to: registration and listing (21 CFR Part 807); labeling (21 CFR Part 801 and Part 809); medical device reporting (reporting of medical device-related adverse events) (21 CFR Part 803) for devices or postmarketing safety reporting (21 CFR Part 4, Subpart B) for combination products (see <https://www.fda.gov/combination-products/guidance-regulatory-information/postmarketing-safety-reporting-combination-products>); good manufacturing practice requirements as set forth in the quality systems (QS) regulation (21 CFR Part 820) for devices or current good manufacturing practices (21 CFR Part 4, Subpart A) for combination products; and, if applicable, the electronic product radiation control provisions (Sections 531-542 of the Act); 21 CFR Parts 1000-1050.

All medical devices, including Class I and unclassified devices and combination product device constituent parts are required to be in compliance with the final Unique Device Identification System rule ("UDI Rule"). The UDI Rule requires, among other things, that a device bear a unique device identifier (UDI) on its label and package (21 CFR 801.20(a)) unless an exception or alternative applies (21 CFR 801.20(b)) and that the dates on the device label be formatted in accordance with 21 CFR 801.18. The UDI Rule (21 CFR 830.300(a) and 830.320(b)) also requires that certain information be submitted to the Global Unique Device Identification Database (GUDID) (21 CFR Part 830 Subpart E). For additional information on these requirements, please see the UDI System webpage at <https://www.fda.gov/medical-devices/device-advice-comprehensive-regulatory-assistance/unique-device-identification-system-udi-system>.

Also, please note the regulation entitled, "Misbranding by reference to premarket notification" (21 CFR 807.97). For questions regarding the reporting of adverse events under the MDR regulation (21 CFR Part 803), please go to <https://www.fda.gov/medical-devices/medical-device-safety/medical-device-reporting-mdr-how-report-medical-device-problems>.

For comprehensive regulatory information about medical devices and radiation-emitting products, including information about labeling regulations, please see Device Advice (<https://www.fda.gov/medical-devices/device-advice-comprehensive-regulatory-assistance>) and CDRH Learn (<https://www.fda.gov/training-and-continuing-education/cdrh-learn>). Additionally, you may contact the Division of Industry and Consumer Education (DICE) to ask a question about a specific regulatory topic. See the DICE website (<https://www.fda.gov/medical-devices/device-advice-comprehensive-regulatory-assistance/contact-us-division-industry-and-consumer-education-dice>) for more information or contact DICE by email (DICE@fda.hhs.gov) or phone (1-800-638-2041 or 301-796-7100).

Sincerely,

Brent Showalter -S

Brent Showalter, Ph.D.

Assistant Director

DHT6B: Division of Spinal Devices

OHT6: Office of Orthopedic Devices

Office of Product Evaluation and Quality

Center for Devices and Radiological Health

Enclosure

Indications for Use

510(k) Number (if known)
K254061

Device Name
Curiteva Porous PEEK Cervical Interbody Fusion System

Indications for Use (Describe)

The Curiteva Porous PEEK Cervical Interbody Fusion System is indicated for intervertebral body fusion of the spine in skeletally mature patients. The Curiteva Porous PEEK Cervical Interbody Fusion System is intended for use for anterior cervical interbody fusion in patients with cervical disc degeneration and/or cervical spinal instability, as confirmed by imaging studies (radiographs, CT, MRI), that results in radiculopathy, myelopathy, and/or pain at multiple contiguous levels from C2 - T1. The Curiteva Porous PEEK Cervical Interbody Fusion System is intended to be used with supplemental fixation. The Curiteva Porous PEEK Cervical Interbody Fusion System is designed for use with autograft, allograft comprised of cortical, cancellous, and/or corticocancellous bone graft, demineralized allograft with bone marrow aspirate, or a bone void filler as cleared by FDA for use in intervertebral body fusion to facilitate fusion.

Type of Use (Select one or both, as applicable)

Prescription Use (Part 21 CFR 801 Subpart D)

Over-The-Counter Use (21 CFR 801 Subpart C)

CONTINUE ON A SEPARATE PAGE IF NEEDED.

This section applies only to requirements of the Paperwork Reduction Act of 1995.

DO NOT SEND YOUR COMPLETED FORM TO THE PRA STAFF EMAIL ADDRESS BELOW.

The burden time for this collection of information is estimated to average 79 hours per response, including the time to review instructions, search existing data sources, gather and maintain the data needed and complete and review the collection of information. Send comments regarding this burden estimate or any other aspect of this information collection, including suggestions for reducing this burden, to:

Department of Health and Human Services
Food and Drug Administration
Office of Chief Information Officer
Paperwork Reduction Act (PRA) Staff
PRASStaff@fda.hhs.gov

"An agency may not conduct or sponsor, and a person is not required to respond to, a collection of information unless it displays a currently valid OMB number."

510(k) Summary**A. Submitter Information**

Submitter: Curiteva, Inc.
 25127 Will McComb Drive
 Tanner, AL 35671
 Phone: (256) 213-1057
 Fax: (256) 213-1058

Contact Person: Eric Linder
regulatory@curiteva.com

Date Prepared: January 14th, 2026

B. Device Information

Trade Name: Curiteva Porous PEEK Cervical Interbody Fusion System

Regulation Number: 21 CFR 888.3080

Regulatory Class: Class II

Product Code(s) &
 Classification Name: Curiteva Porous PEEK Cervical Interbody Fusion System
 ODP – Intervertebral fusion Device with Bone Graft, Cervical
 (21 CFR Part §888.3080)

Classification Panel: Division of Orthopedic Devices

Predicate Device: Curiteva Porous PEEK Cervical Interbody System (K213030)

Reference Devices: Alphatec IdentiTi NanoTec Cervical Interbody System
 (K241375)
 Globus Patriot® Colonial® Spacer (K143578)
 Curiteva Cervical Interbody System (K181261)

C. Device Description

The Curiteva Porous PEEK Cervical Interbody Fusion System implants are sterile, single-use devices and available in a variety of different footprints, styles and sizes to accommodate the individual pathology and anatomical conditions of the patient. The implants are generally box-shaped with an open central corridor to permit packing with bone graft to facilitate fusion. The implants have a dense central ring with a porous structure lining the vertical graft corridor and on the superior and inferior surfaces of the construct. Each implant has been surface treated with a hydroxyapatite (HA) coating that is approximately 20nm thick.

The Curiteva Porous PEEK Cervical Interbody Fusion System implants are manufactured from implant-grade PEEK (per ASTM F2026) with Titanium alloy markers (per ASTM F136).

The purpose of this submission is to update the indications for use and to introduce new sizes to the Curiteva Porous PEEK Cervical Interbody Fusion System.

D. Indications for Use

The Curiteva Porous PEEK Cervical Interbody Fusion System is indicated for intervertebral body fusion of the spine in skeletally mature patients. The Curiteva Porous PEEK Cervical Interbody Fusion System is intended for use for anterior cervical interbody fusion in patients with cervical disc degeneration and/or cervical spinal instability, as confirmed by imaging studies (radiographs, CT, MRI), that results in radiculopathy, myelopathy, and/or pain at multiple contiguous levels from C2 - T1. The System is intended to be used with supplemental fixation. The system is designed for use with autograft, allograft comprised of cortical, cancellous, and/or corticocancellous bone graft, demineralized allograft with bone marrow aspirate, or a bone void filler as cleared by FDA for use in intervertebral body fusion to facilitate fusion.

E. Comparison of Technological Characteristics

The technological design features of the subject implants were compared to the predicates in intended use, indications for use, design, function and technology and it was demonstrated that they are substantially equivalent.

F. Performance Data

Performance data is not included in this submission.

G. Conclusion

Based on the indications for use, technological characteristics, non-clinical performance testing, and comparison to predicate devices, the subject devices have shown to be substantially equivalent to legally marketed predicate devices.